

Application No. 10/602,128
Amendment, dated January 26, 2005
Reply to Office Action of November 26, 2004

AMENDMENTS TO THE CLAIMS

Please amend claims 4, 5, 12 and 13 as follows:

1. (Previously Presented) A magnification loupe carried by a user wearable device, comprising:

 a housing having a first end with a first aperture for supporting an eyepiece lens, and a second end with a second aperture for supporting an objective lens;

 an eyepiece lens disposed in said first end of said housing; and

 an objective lens disposed in said second end of said housing;

 said objective lens having a non-circular shape, wherein at least two oppositely disposed arcuate first peripheral edges are defined by a first radius extending from a first center, and wherein at least two oppositely disposed arcuate second peripheral edges are defined by at least one second radius extending from at least a second center not coincident with said first center, said second radius having a length different from said first radius.

2. (Original) The magnification loupe of claim 1, wherein said second peripheral edges are semi-elliptical in shape.

3. (Original) The magnification viewer of claim 1, wherein said eyepiece lens comprises a single lens element, and wherein said objective lens comprises two lens elements.

Application No. 10/602,128
 Amendment. dated January 26, 2005
 Reply to Office Action of November 26, 2004

4.(Currently Amended) The magnification loupe of claim 3, wherein said eyepiece lens and said objective lens are constructed and arranged according to the following parameters:

Element	Glass	n_d	n_d	Radius	Thickness	Maximum Diameter	Sep.
I	Schott NSK5	1.589	61.3	$R_1=\infty$ $R_2=\infty$	2.2	12.0	
II	Schott NBK7	1.517	64.2	$R_3=36.49$ $R_4=18.48$	1.5	12.0	$S_1=0.6$
III	Schott NSF56	1.805	25.4	$R_5=85.68$ $R_6=39.71$	1.6	$D_1=22.24$ $D_2=23.60$	$S_2=14.46$
IV	Schott NBK7	1.517	64.2	$R_7=39.71$ $R_8=21.55$	6.65	$D_3=23.60$ $D_4=23.60$	

wherein the radius, thickness, and separation dimensions are given in millimeters; Roman numerals identify the lens elements in their respective order from the eyepoint side to the object side and element I is a representative correction lens of the user wearable device; n_d represents the refractive index of each element; n_d is the abbe dispersion number; R_1 , R_2 , etc. represent the radii of the respective refractive surfaces in order, from the eyepoint side to the object side; D_1 , and D_2 , etc. represent the maximum

Application No. 10/602,128
 Amendment. dated January 26, 2005
 Reply to Office Action of November 26, 2004

clear lens aperture diameters of the parent lens elements; and S_1 , S_2 represent the air space between the elements, measured along an optical centerline.

5. (Currently Amended) The magnification loupe of claim 3, wherein said eyepiece lens and said objective lens are constructed and arranged according to the following parameters:

Element	Glass	h_d	n_d	Radius	Thickness	Maximum Diameter	Sep.
I	Schott NSK5	1.589	61.3	$R_1=98.19$ $R_2=98.19$	3.0	25.4	
II	Schott NBALF4	1.580	53.9	$R_3=52.10$ $R_4=20.16$	1.5	$D_1=13.00$ $D_2=13.25$	$S_1=4.1$
III	O'Hara STIH23	1.785	26.3	$R_5=85.68$ $R_6=43.17$	1.8	26.15	$S_2=13.59$
IV	Schott NBK7	1.517	64.2	$R_7=43.17$ $R_8=22.39$	7.6	26.15	

wherein the radius, thickness, and separation dimensions are given in millimeters;
 Roman numerals identify the lens elements in their respective order from the eyepoint side to the object side and element I is a representative correction lens of the user wearable device; h_d represents the refractive index of each element; n_d is the abbe dispersion number; R_1 , R_2 , etc. represent the radii of the respective refractive surfaces in order, from the eyepoint side to the object side; D_1 , and D_2 , [[etc.]] represent the maximum clear lens aperture diameters of the parent lens elements; and S_1 , S_2 represent the air space between the elements, measured along an optical centerline.

Application No. 10/602,128
Amendment. dated January 26, 2005
Reply to Office Action of November 26, 2004

6. (Original) The magnification loupe of claim 1, further comprising a correction lens couplable to said housing proximate said first end.

7. (Previously Presented) The magnification loupe of claim 6, wherein said correction lens is interchangeably coupled to said housing, whereby said correction lens is selectively removed from said housing and replaced with a different correction lens to thereby vary a working distance of the loupe.

8. (Previously Presented) The magnification loupe of claim 1, wherein said housing is configured for mounting through a lens of the user wearable device.

9. (Previously Presented) The magnification loupe of claim 1, wherein said housing is configured to be mounted to the user wearable device by a mounting member secured to a frame of the user wearable device.

10. (Previously Presented) A magnification viewer, comprising:

a user wearable device having a frame and at least one eyeglass lens supported on said frame; and

at least one magnification loupe operatively coupled to said user wearable device, said magnification loupe comprising:

a housing having a first end with a first aperture for supporting an eyepiece lens, and a second end with a second aperture for supporting an objective lens,

Application No. 10/602,128
Amendment, dated January 26, 2005
Reply to Office Action of November 26, 2004

an eyepiece lens disposed in said first end of said housing, and
an objective lens disposed in said second end of said housing,
said objective lens having a non-circular shape, wherein at least two
oppositely disposed arcuate first peripheral edges are defined by a first radius extending
from a first center, and wherein at least two oppositely disposed arcuate second
peripheral edges are defined by at least one second radius extending from at least a
second center not coincident with said first center, said second radius having a length
different from said first radius.

11. (Previously Presented) The magnification viewer of claim 10, further comprising a
correction lens interchangeably coupled to said housing proximate said first end, whereby
said correction lens is selectively removed from said housing and replaced with a
different correction lens to thereby vary a working distance of said loupe.

12. (Currently Amended) A magnification loupe carried by a user wearable device,
comprising:

a housing having a first end with a first aperture for supporting an eyepiece
lens, and a second end with a second aperture for supporting an objective lens;

a single element eyepiece lens disposed in said first end of said housing;

and

a two element objective lens disposed in said second end of said housing;

Application No. 10/602,128
 Amendment. dated January 26, 2005
 Reply to Office Action of November 26, 2004

said eyepiece lens and said objective lens constructed and arranged
 according to the following parameters:

Element	Glass	n_d	n_d	Radius	Thickness	Maximum Diameter	Sep.
I	Schott NSK5	1.589	61.3	$R_1 = [[00]]$ ∞ $R_2 = [[00]]$ ∞	2.2	12.0	
II	Schott NBK7	1.517	64.2	$R_3 = 36.49$ $R_4 = 18.48$	1.5	12.0	$S_1 = 0.6$
III	Schott NSF56	1.805	25.4	$R_5 = 85.68$ $R_6 = 39.71$	1.6	$D_1 = 22.24$ $D_2 = 23.60$	$S_2 = 14.46$ [[6]]
IV	Schott NBK7	1.517	64.2	$R_7 = 39.71$ $R_8 = 21.55$	6.65	$D_3 = 23.60$ $D_4 = 23.60$	

wherein the radius, thickness, and separation dimensions are given in millimeters;
 Roman numerals identify the lens elements in their respective order from the eyepoint
 side to the object side and element I is a representative correction lens of the user
~~wearable device~~; n_d represents the refractive index of each element; n_d is the abbe
 dispersion number; R_1 , R_2 , etc. represent the radii of the respective refractive surfaces in
 order, from the eyepoint side to the object side; D_1 and D_2 , etc. represent the maximum
clear lens aperture diameters of the parent lens elements; and S_1 , S_2 represent the air
 space between the elements, measured along an optical centerline.

Application No. 10/602,128
 Amendment. dated January 26, 2005
 Reply to Office Action of November 26, 2004

13. (Currently Amended) A magnification loupe carried by a user wearable device, comprising:

a housing having a first end with a first aperture for supporting an eyepiece lens, and a second end with a second aperture for supporting an objective lens;

a single element eyepiece lens disposed in said first end of said housing;
 and

a two element objective lens disposed in said second end of said housing;

said eyepiece lens and said objective lens constructed and arranged according to the following parameters:

Element	Glass	h_d	n_d	Radius	Thickness	Maximum Diameter	Sep.
I	Schott NSK5	1.589	61.3	$R_1=98.19$ $R_2=98.19$	3.0	25.4	
II	Schott NBALF4	1.580	53.9	$R_3=52.10$ $R_4=20.16$	1.5	$D_1=13.00[[D]]$ $D_2=13.25$	$S_1=4.1$
III	O'Hara STIH23	1.785	26.3	$R_5=85.68$ $R_6=43.17$	1.8	26.15	$S_2=13.59$
IV	Schott NBK7	1.517	64.2	$R_7=43.17$ $R_8=22.39$	7.6	26.15	

wherein the radius, thickness, and separation dimensions are given in millimeters;
 Roman numerals identify the lens elements in their respective order from the eyepoint side to the object side; h_d represents the refractive index of each element; n_d is the abbe dispersion number; R_1 , R_2 , etc. represent the radii of the respective refractive surfaces in order, from the eyepoint side to the object side and element I is a representative correction lens of the user wearable device; D_1 and D_2 , etc. represent the maximum clear lens aperture diameters of the parent lens elements; and S_1 , S_2 represent the air space

Application No. 10/602,128
Amendment, dated January 26, 2005
Reply to Office Action of November 26, 2004

between the elements, measured along an optical centerline.